KERN & Sohn GmbH
Ziegelei 1
D-72336 Balingen
E-Mail: info@kern-sohn.com
Tel: +49-[0]7433- 9933-0
Fax: +49-[0]7433-9933-149
Internet: www.kern-sohn.com

Service manual
Mechanical height-measuring rod

KERN MSB 80
Version 1.0
3/2010
GB



MSB-SH-e-1010

GB

KERN MSB 80

Version 1.0 3/2010

Service manual

Mechanical height-measuring rod

Table of Contents

1 Basic Information 3

2 Technical Data 3

3 Assembly Drawing 4

1 Basic Information

Grundlegende Hinweise

The device must be repaired only by trained specialist staff or personnel with professional formation (such as a repair-specialist accredited by law concerning verification).

The service manual is obligatory for repair work.

After repair, original conditions of the device have to be restored.

Only original spare parts should be used.

Das Gerät darf nur von geschultem oder beruflich ausgebildetem Fachpersonal (z. B. eichrechtlich anerkannter Instandsetzer) repariert werden.

Die Serviceanleitung ist bindend für Reparaturen.

Das Gerät muss nach erfolgter Reparatur wieder in den Originalzustand zurückversetzt werden.

Es dürfen nur Originalersatzteile verwendet werden.

2 Technical Data

KERN	**MSB 80**
Measuring range	10 – 80 cm
Readability	1 mm
tolerance	5 mm
Dimensions (w x h x d) mm	890 x 100 x 330 mm
Temperature range	+10°C … +40°C
Storage temperature	- 20°C … + 60°C
Admitted as medical product as per Guideline 93/42/EWG (CE 0297)	Category I with measuring function

3 Assembly Drawing

MSB 80 (HM-80P)

