

EU-Konformitätserklärung | EU Declaration of Conformity

DE Wir erklären hiermit unter alleiniger Verantwortung, dass das Produkt, auf das sich diese Erklärung bezieht, mit den nachstehenden Richtlinien übereinstimmt. Das Produkt erfüllt die einschlägigen Harmonisierungsrechtsvorschriften der Union. Für die Waage liegt eine Bauartzulassung vor. Das Jahr der EU-Konformitätsbewertung ist neben der CE Konformitätskennzeichnung aufgeführt. Solche Waagen sind ab Werk EU-konformitätsbewertet und tragen die Kennzeichnung „M“ auf dem Gerät selbst und auf der Verpackung. Die angegebene Gravitationszone legt den Verwendungsort fest. Ein Wechsel des Gebrauchsortes über die Grenzen des angegebenen Verwendungsbereiches hinaus macht eine erneute Eichung erforderlich.

EN We hereby declare and assume sole responsibility for the declaration that the product complies with the directives hereinafter. The object of the declaration described below is in conformity with the relevant Union harmonisation legislation. The scale was issued with a type approval certificate. The year of EU conformity assessment can be found next to the CE mark. Such scales bear the mark "M", on the device itself as well as on the packaging. The gravity zone specified determines the place of use.

Typ | Type |

Seriennr. | Serial no. |

CPB 15K2DM

XXXXXXXXXX

CPB 30K5DM

CPB 6K1DM

CE Kennzeichnung CE mark applied	EU-Richtlinie EU directive	Normen Standards	Bauartzulassungen Type approvals
CE	2011/65/EU (RoHS)	EN 63000:2018	
CE	2014/30/EU (EMC)	EN 55032:2015+A11:2020 EN 55035:2017 EN 61000-3-3:2013+A1:2019 EN IEC 61000-3-2:2019	
CE M21 0122¹⁾	2014/31/EU (NAWI)	EN 45501:2015	0200-NAWI-02699
CE	2014/35/EU (LVD)	EN 62368-1:2014+A11:2017	

¹⁾ Diese CE Kennzeichnung kennzeichnet Konformitätsbewertung durch KERN; Eignung für Anwendungsbereiche nach 2014/31/EU, Kapitel 1. Artikel 1 Pkt. 2 (a bis f). Diese Waagen tragen das Metrologiekennzeichen „M“ gefolgt von der Jahreszahl der EU-Konformitätsbewertung auf dem Gerät. Für die Waage liegt eine EU-Baumusterprüfung nach 2014/31/EU vor. Die angegebene Gravitationszone legt den Verwendungsort fest. Ein Wechsel des Gebrauchsortes über die Grenzen des angegebenen Verwendungsbereiches hinaus macht eine erneute Prüfung erforderlich. Die benannte Stelle NMi Certin BV (0122) führt Audits für Modul D gemäß Richtlinie 2014/31/EU durch und stellt das Zertifikat CE-240 für KERN aus.

¹⁾ This CE mark applied indicates declaration of conformity by KERN; approved for categories of use as listed in 2014/31/EU, chapter 1. article 1 pt. 2 (a to f). These weighing instruments bear the metrology marking "M" followed by the last two digits of the year of the declaration of conformity. Associated there is a type examination certificate according to 2014/31/EU. The specified gravitational zone determines the place of use. A change beyond the limits of the specified area requires a new verification. The notified body NMi Certin BV (0122) carries out audits for module D according to directive 2014/31/EU and issued the certificate CE-240 for KERN.

g = ...

Ort oder Zone: ...

Location or zone:

Datum | Date |: 05.03.2021

Ort der Ausstellung: 72336 Balingen,
Place of issue: Germany

John Doe
KERN & Sohn GmbH

Albert Sauter
KERN & Sohn GmbH

Signatur:
Signature:

Prüfbevollmächtigter
Verification officer

Geschäftsführer
Managing director