

## PRODUCT BROCHURE

---

# KERN 315-309-205

Date: 19.03.2025

ENGLISH



Safety Set 500 g (F2); 20 g (F1)

---

## CONTACT

KERN & SOHN GmbH  
Ziegelei 1  
72336 Balingen  
Germany

Phone : +49 7433 9933-0  
Fax : +49 7433 9933-149  
Email : [info@kern-sohn.com](mailto:info@kern-sohn.com)  
Web : [www.kern-sohn.com](http://www.kern-sohn.com)

## PRODUCT SHOWCASE



## DESCRIPTION

- These “KERN Safety Sets” each contain one test weight for testing the sensitivity, i.e. the correct adjustment of your balance as well as a small test weight to test the lower end of the weighing range, the so-called minimum sample weight
- As an option, the “KERN Safety Set” has space for another test weight, for testing your balance at a weight which is relevant for you
- Useful accessories which have been selected to suit that particular “KERN Safety Set”, such as, for example, special gloves, tweezers, weight grips, brushes, etc. will assist you in handling your test weights properly
- Stored in the practical protective case next to your balance, you can check and ensure the high precision of your balance at any time

## SPECIFICATIONS

### Article codes, product groups & Status information

Model article-code	315-309-205
Model-Serie	315
Assortment	KERN
Product Type	Test weights
Product group	Safety sets
Product usage type	Main item;Accessories/Services

EAN code (Model) 4045761223417

Customs tariff number 84239090

### Technical data - Measuring system

Nominal weights, included 500 g (F2); 20 g (F1)

Nominal value 20 g & 500 g

OIML class F1;F2

### Technical data - Packing & Shipping

Dimensions packaging (W×D×H)	255×135×140 mm
Gross weight	3 kg
Net weight	1.2 kg
Shipping method	Parcel service
Packaging component - by weight - plastics	1.8 kg
Delivery time	1 d

## Technical data - Verification

Verification possible	yes
-----------------------	-----

## Optional Services

DAkkS Calibration	962-400-400
-------------------	-------------

## FUNCTIONS

Standard
----------



## SERVICES

Model	Description
921-439	Guaranteed Plus-deviation (? +1/3 OIML) KERN 921-439
962-400-400	DAkkS Calibration KERN 962-400-400

## SPARE PARTS

Model	Description
315-242	Tweezers KERN 315-242